

# EC CERTIFICATE

## Full Quality Assurance System

Certificate No.:  
215934-2017-CE-KOR-NA-PS Rev 4.0

Project No.:  
PRJC-533956-2015-MSL-KOR

Valid Until:  
30 May 2021

This is to certify that the quality system of:

### **Bistos Co., Ltd.**

7th Fl., A Bldg., Woolim Lions Valley 5-cha, 302, Galmachi-ro, Jungwon-gu,  
Seongnam-si, Gyeonggi-do 13201, Korea

For design, production and final product inspection/testing of:

### **ELECTRIC BREAST PUMP, INFANT WARMER, INFANT INCUBATOR**

Has been assessed with respect to:

### **THE CONFORMITY ASSESSMENT PROCEDURE DESCRIBED IN ANNEX II EXCLUDING SECTION 4 OF COUNCIL DIRECTIVE 93/42/EEC ON MEDICAL DEVICES, AS AMENDED**

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and date:  
**Høvik, 31 March 2020**



For:  
**DNV GL PRESAFE AS**  
**Notified Body No.: 2460**

**Mariann Jeremiassen**

The certificate is digitally verified by blockchain technology. For more info, see [www.dnvgl.com/assurance/certificates-in-the-blockchain.html](http://www.dnvgl.com/assurance/certificates-in-the-blockchain.html)



Notice: The Certificate is subject to terms and conditions as set out in the Certification Agreement. Failure to comply may render this Certificate invalid.  
NOTIFIED BODY: DNV GL PRESAFE AS, Veritasveien 3, N-1363 Høvik, Norway - Registered Enterprise No: NO 997 067 401 MVA .

Certificate No.:  
215934-2017-CE-KOR-NA-PS Rev 4.0

Project No.:  
PRJC-533956-2015-MSL-KOR

Valid Until:  
30 May 2021

## Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as "Forskrift om Medisinsk Utstyr" by the Norwegian Ministry of Health and Care Services.

Certificate history:

Revision	Description	Issue Date
0.0	Original Certificate	18 April 2017
1.0	New model (Milk Genie) added	25 September 2017
2.0	Product add (Infant incubator)	03 January 2018
3.0	EU Rep change	13 April 2018
<b>4.0</b>	<b>New model (BT-150, BT-150L, BT-150B, BT-150S) added (In bold)</b>	<b>31 March 2020</b>

Products covered by this Certificate:

Product Description	Product Name	Class
Electric Breast Pump	<ul style="list-style-type: none"> <li>▪ Hi bebe<sup>plus</sup> BT-100</li> <li>▪ Milk Genie</li> <li>▪ <b>BT-150, BT-150L, BT-150B, BT-150S</b></li> </ul>	IIa
Infant warmer	<ul style="list-style-type: none"> <li>▪ BT-550</li> </ul>	IIb
Infant incubator	<ul style="list-style-type: none"> <li>▪ BT-500</li> </ul>	IIb

The complete list of devices is filed with the Notified Body

## Sites covered by this certificate

Site Name	Address
Bistos Co., Ltd.	7th Fl., A Bldg., Woolim Lions Valley 5-cha, 302, Galmachi-ro, Jungwon-gu, Seongnam-si, Gyeonggi-do 13201, Korea

## EU Representative

OBELIS S.A, Bd. General Wahis, 53, 1030 Brussels, Belgium

Certificate No.:  
215934-2017-CE-KOR-NA-PS Rev 4.0

Project No.:  
PRJC-533956-2015-MSL-KOR

Valid Until:  
30 May 2021

## Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

## Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate